Fellows Research Centre
North Dean Business Park,
Stainland Road, Greetland,
Halifax, HX4 8LR

T +44 (0)20 7993 0030
E info@gamahealthcare.com
gamahealthcare.com

SEERS Medical and GAMA Healthcare Ltd Compatibility statement – Clinell Universal

GAMA Healthcare Ltd is pleased to confirm the successful completion of material compatibility testing conducted in collaboration with Pocketalk. The evaluation demonstrated full compatibility between our Clinell Universal Surface Range (including CW200 and CW200B) and the SEERS Medical devices with all materials performing within the specified parameters and exhibiting no material degradation or adverse effects during and after the two-week of continuous immersion and a 10,000-wipe scrub test.

These results validate the suitability of the Clinell Universal Surface Range for use with the SEERS Medical devices and suggest that there is a low risk of adverse effects which can be caused by cleaning the device routinely with the Clinell Universal Surface Range. Based on this both parties consider the device to be compatible with the Clinell Universal Surface Range produced by GAMA Healthcare Ltd.

Author: Ryan Weller, Materials Scientist, r.weller@gamahealthcare.com

Signed for and on behalf of

Gama Healthcare Ltd

Signature: 

Name: Ryan Weller

Position: Materials Scientist

Date: 11/11/2025

Signed for and on behalf

SEERS Medical

Signature: 

Name: Dann Rees

Position: Head of Sales & Marketing

Date: 11/11/2025